

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 006282  
15:12 11/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:09  
STERILIZING 4:11  
136.5 C 223kPa 4:11  
136.4 C 225kPa 4:41  
136.4 C 224kPa 5:11  
136.5 C 224kPa 5:41  
136.4 C 224kPa 6:11  
136.4 C 223kPa 6:41  
136.2 C 223kPa 7:11  
136.2 C 223kPa 7:41  
136.2 C 223kPa 7:41  
VENTING 7:41  
AIR DRYING 8:36  
CYCLE COMPLETE 8:50

-----  
Digital Signature #  
6CAD7527CEFFE433  
-----