
Midmark M11 - v1.0.5
Total Cycles: 3056

Name: RITTER2
Sterilizer ID: V2313228
10/27/2023 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 148.8 F 0.8
4:00 154.0 F 3.8
6:00 180.5 F 8.3
8:00 204.5 F 13.2
10:00 218.7 F 7.6
12:00 232.3 F 12.9
14:00 235.2 F 10.1
16:00 235.3 F 8.5
18:00 249.0 F 15.4
20:00 261.3 F 22.4
22:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 271.9 F 29.2
2:00 272.4 F 29.5
3:00 272.3 F 29.9
4:00 271.5 F 29.1

Min 271.5 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 203.2 F 0.0
10:00 206.9 F 0.0
15:00 199.7 F 0.0
20:00 193.9 F 0.0
25:00 189.7 F 0.0
30:00 187.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

10/27/2023 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____