
Midmark M11 - v1.0.0
Total Cycles: 3094

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2023 03:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 217.9 F 0.0
2:00 209.6 F 4.3
4:00 223.4 F 10.0
6:00 228.4 F 6.3
8:00 238.1 F 11.7
10:00 237.0 F 9.5
12:00 238.8 F 8.9
14:00 250.1 F 15.4
16:00 262.5 F 22.5
18:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 29.0
2:00 271.9 F 28.8
3:00 272.4 F 28.9
4:00 271.6 F 28.6

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 227.9 F 0.0
10:00 241.2 F 0.0
15:00 239.0 F 0.0
20:00 231.5 F 0.0
25:00 222.6 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 2:48
HEATING: 18:11
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

04/11/2023 04:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____