

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1114

Name: RITTER 1  
Sterilizer ID: V2178198  
08/24/2021 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
2:00 180.5 F 2.3  
4:00 194.6 F 6.3  
6:00 223.4 F 11.8  
8:00 225.4 F 6.4  
10:00 236.4 F 11.5  
12:00 234.8 F 8.7  
14:00 241.3 F 10.6  
16:00 246.4 F 13.4  
18:00 258.2 F 19.8  
20:00 268.8 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.4 F 28.9  
2:00 271.4 F 28.5  
3:00 272.4 F 28.9  
4:00 271.3 F 28.3

Min 271.2 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 221.5 F 0.0  
10:00 224.3 F 0.0  
15:00 226.5 F 0.0  
20:00 226.7 F 0.0  
25:00 210.1 F 0.0  
30:00 195.1 F 0.0  
Drying Complete

FILLING: 3:07  
HEATING: 21:10  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:32  
CYCLE COMPLETE

08/24/2021 10:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_