
Midmark M11 - v1.0.5
Total Cycles: 2585

Name: RITTER2
Sterilizer ID: V2313228
06/29/2023 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.3 F 0.0
2:00 135.9 F 0.0
4:00 141.0 F 0.1
6:00 168.9 F 3.8
8:00 188.1 F 8.2
10:00 212.5 F 13.1
12:00 222.0 F 7.6
14:00 234.7 F 12.6
16:00 235.3 F 9.7
18:00 237.4 F 9.3
20:00 247.4 F 14.4
22:00 259.3 F 21.0
24:00 270.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.8 F 29.8
2:00 272.0 F 29.2
3:00 271.5 F 29.2
4:00 272.7 F 29.8

Min 271.3 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 203.2 F 0.0
10:00 208.5 F 0.0
15:00 209.7 F 0.0
20:00 209.9 F 0.0
25:00 211.7 F 0.0
30:00 210.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 24:56
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

06/29/2023 10:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____