
Midmark M11 - v1.0.5
Total Cycles: 3505

Name: RITTER2
Sterilizer ID: V2313228
04/08/2024 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.3 F 0.0
2:00 96.0 F 0.0
4:00 120.0 F 0.2
6:00 149.0 F 1.3
8:00 200.2 F 4.2
10:00 210.7 F 7.5
12:00 216.6 F 11.2
14:00 220.0 F 5.2
16:00 228.1 F 8.6
18:00 238.5 F 13.0
20:00 238.1 F 10.1
22:00 237.6 F 9.2
24:00 247.7 F 14.3
26:00 258.6 F 20.3
28:00 268.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.4 F 29.5
2:00 271.8 F 28.8
3:00 271.4 F 29.1
4:00 272.9 F 30.0

Min 271.0 F 28.6
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 228.4 F 0.0
10:00 219.9 F 0.0
15:00 194.5 F 0.0
20:00 189.7 F 0.0
25:00 185.4 F 0.0
30:00 182.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:14
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:43
CYCLE COMPLETE

04/08/2024 03:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____