

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002661  
15:55 07/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:16  
STERILIZING 4:18  
136.0 C 220kPa 4:18  
136.2 C 223kPa 4:48  
136.1 C 223kPa 5:18  
136.4 C 225kPa 5:48  
136.4 C 226kPa 6:18  
136.3 C 225kPa 6:48  
136.4 C 225kPa 7:18  
136.3 C 225kPa 7:48  
136.3 C 225kPa 7:48  
VENTING 7:48  
AIR DRYING 8:47  
CYCLE COMPLETE 11:35

-----  
Digital Signature #  
D764FA6C10CA61EC  
-----