

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2111

Name: RITTER2  
Sterilizer ID: V2313228  
03/01/2023 04:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.9 F 0.0  
2:00 140.0 F 2.1  
4:00 154.3 F 5.3  
6:00 185.8 F 9.8  
8:00 215.9 F 6.1  
10:00 228.2 F 9.2  
12:00 233.3 F 9.6  
14:00 238.2 F 12.0  
16:00 238.7 F 10.5  
18:00 251.9 F 17.0  
20:00 263.4 F 23.9  
22:00 272.1 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.4  
1:00 271.3 F 29.7  
2:00 272.4 F 30.2  
3:00 272.2 F 29.8  
4:00 272.2 F 30.0

Min 271.0 F 29.2  
Max 273.6 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
5:00 229.1 F 0.0  
10:00 229.6 F 0.0  
15:00 202.1 F 0.0  
20:00 196.4 F 0.0  
25:00 205.9 F 0.0  
30:00 215.3 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 22:03  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:27  
CYCLE COMPLETE

03/01/2023 05:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_