

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4049

Name: RITTER 1  
Sterilizer ID: V2178198  
11/22/2023 08:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.5 F 0.0  
2:00 82.6 F 0.0  
4:00 111.7 F 0.4  
6:00 122.5 F 2.6  
8:00 145.4 F 5.0  
10:00 163.8 F 7.8  
12:00 183.4 F 11.1  
14:00 211.1 F 5.3  
16:00 212.2 F 5.4  
18:00 221.9 F 8.8  
20:00 231.7 F 12.6  
22:00 231.9 F 7.8  
24:00 243.6 F 12.6  
26:00 240.1 F 10.4  
28:00 252.0 F 16.3  
30:00 262.6 F 22.3  
32:00 272.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.7 F 28.9  
2:00 271.4 F 28.1  
3:00 272.7 F 29.0  
4:00 271.9 F 28.3  
  
Min 271.3 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.3 F 0.0  
5:00 200.0 F 0.0  
10:00 198.2 F 0.0  
15:00 197.5 F 0.0  
20:00 221.3 F 0.0  
25:00 230.9 F 0.0  
30:00 229.2 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:30  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:10:55  
CYCLE COMPLETE

11/22/2023 10:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_