
Midmark M11 - v1.0.5
Total Cycles: 1164

Name: RITTER2
Sterilizer ID: V2313228
06/07/2022 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.5 F 0.0
2:00 149.7 F 0.9
4:00 154.1 F 4.0
6:00 179.0 F 8.4
8:00 207.6 F 13.5
10:00 221.0 F 8.2
12:00 235.1 F 13.5
14:00 236.0 F 10.5
16:00 236.4 F 8.5
18:00 248.5 F 15.3
20:00 260.1 F 21.8
22:00 270.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.4 F 30.0
2:00 271.3 F 29.3
3:00 272.2 F 30.2
4:00 272.0 F 29.7

Min 271.2 F 29.3
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 214.6 F 0.0
10:00 224.0 F 0.0
15:00 227.4 F 0.0
20:00 224.2 F 0.0
25:00 223.7 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

06/07/2022 09:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____