

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3780

Name: RITTER 1  
Sterilizer ID: V2178198  
09/15/2023 10:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.4 F 0.0  
2:00 149.8 F 0.0  
4:00 167.5 F 1.3  
6:00 188.8 F 4.1  
8:00 219.1 F 7.6  
10:00 222.6 F 11.6  
12:00 226.2 F 5.8  
14:00 237.1 F 10.5  
16:00 234.4 F 8.0  
18:00 245.4 F 13.0  
20:00 243.3 F 11.6  
22:00 255.4 F 17.8  
24:00 266.0 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 27.9  
1:00 272.5 F 28.5  
2:00 271.2 F 27.8  
3:00 272.3 F 28.6  
4:00 271.6 F 28.0

Min 271.2 F 27.8  
Max 272.5 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
5:00 232.8 F 0.0  
10:00 223.8 F 0.0  
15:00 222.5 F 0.0  
20:00 213.7 F 0.0  
25:00 215.5 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 25:44  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:04:12  
CYCLE COMPLETE

09/15/2023 11:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_