
Midmark M11 - v1.0.0
Total Cycles: 1000

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2021 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 178.9 F 2.0
4:00 180.5 F 5.7
6:00 211.1 F 10.8
8:00 220.2 F 5.6
10:00 232.3 F 11.1
12:00 235.6 F 8.5
14:00 241.8 F 10.2
16:00 246.6 F 13.9
18:00 259.0 F 20.7
20:00 270.4 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.7 F 29.3
2:00 271.3 F 28.7
3:00 272.9 F 29.3
4:00 271.6 F 28.7

Min 271.3 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 217.7 F 0.0
10:00 232.5 F 0.0
15:00 232.7 F 0.0
20:00 230.1 F 0.0
25:00 228.4 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

07/13/2021 08:54 AM
ACCEPTED: X
REJECTED:
Operator: U01