
Midmark M11 - v1.0.5
Total Cycles: 1241

Name: RITTER2
Sterilizer ID: V2313228
06/29/2022 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.9 F 0.0
2:00 117.8 F 1.1
4:00 138.4 F 4.2
6:00 164.4 F 7.9
8:00 191.6 F 12.3
10:00 208.6 F 5.8
12:00 222.6 F 10.5
14:00 227.9 F 7.3
16:00 238.2 F 12.1
18:00 238.1 F 10.2
20:00 250.7 F 16.4
22:00 261.7 F 22.8
24:00 272.3 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.5 F 30.1
2:00 272.1 F 29.5
3:00 271.4 F 29.4
4:00 272.5 F 30.1

Min 271.3 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 203.4 F 0.0
10:00 215.6 F 0.0
15:00 222.8 F 0.0
20:00 223.5 F 0.0
25:00 224.1 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:41
CYCLE COMPLETE

06/29/2022 12:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____