
Midmark M11 - v1.0.5
Total Cycles: 2610

Name: RITTER2
Sterilizer ID: V2313228
07/05/2023 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.8 F 0.0
2:00 112.0 F 0.0
4:00 128.5 F 0.4
6:00 170.5 F 4.3
8:00 191.8 F 8.1
10:00 211.4 F 12.5
12:00 221.7 F 7.1
14:00 233.5 F 12.0
16:00 234.3 F 9.1
18:00 239.3 F 10.2
20:00 246.2 F 13.8
22:00 258.5 F 20.5
24:00 269.4 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.5
2:00 271.4 F 29.1
3:00 272.5 F 29.6
4:00 271.4 F 28.8

Min 271.3 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 206.1 F 0.0
10:00 220.2 F 0.0
15:00 227.4 F 0.0
20:00 223.3 F 0.0
25:00 220.2 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:03:08
CYCLE COMPLETE

07/05/2023 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____