
Midmark M11 - v1.0.5
Total Cycles: 237

Name: RITTER2
Sterilizer ID: V2313228
07/22/2021 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.5 F 0.0
2:00 127.5 F 1.5
4:00 157.4 F 4.9
6:00 182.6 F 9.1
8:00 204.5 F 7.8
10:00 226.0 F 7.7
12:00 234.3 F 12.7
14:00 237.1 F 9.9
16:00 236.8 F 9.0
18:00 248.7 F 14.9
20:00 260.4 F 21.4
22:00 271.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.3 F 29.0
2:00 272.2 F 29.5
3:00 271.6 F 28.8
4:00 272.0 F 29.5

Min 271.1 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.7 F 0.0
5:00 222.4 F 0.0
10:00 193.3 F 0.0
15:00 192.3 F 0.0
20:00 189.5 F 0.0
25:00 186.7 F 0.0
30:00 182.8 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:14
CYCLE COMPLETE

07/22/2021 10:56 AM
ACCEPTED: X
REJECTED:
Operator: U02