
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002208
11:58 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:35
STERILIZING 3:47
136.4 C 224kPa 3:47
136.3 C 225kPa 4:17
136.4 C 227kPa 4:47
136.4 C 226kPa 5:17
136.2 C 223kPa 5:47
136.1 C 222kPa 6:17
136.3 C 225kPa 6:47
136.3 C 225kPa 7:17
136.3 C 225kPa 7:17
VENTING 7:17
AIR DRYING 8:10
CYCLE COMPLETE 16:02

Digital Signature #
BF955A709C29F1C1
