
Midmark M11 - v1.0.5
Total Cycles: 884

Name: RITTER2
Sterilizer ID: V2313228
03/16/2022 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 182.9 F 3.4
4:00 186.8 F 8.0
6:00 223.5 F 11.3
8:00 230.4 F 9.7
10:00 234.5 F 8.7
12:00 243.5 F 12.6
14:00 244.4 F 13.0
16:00 257.8 F 20.1
18:00 269.7 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.4 F 30.1
2:00 272.4 F 29.6
3:00 271.5 F 29.1
4:00 271.5 F 29.7

Min 271.2 F 29.1
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 234.3 F 0.0
10:00 229.7 F 0.0
15:00 223.4 F 0.0
20:00 221.9 F 0.0
25:00 221.7 F 0.0
30:00 223.6 F 0.0

Drying Complete

FILLING: 1:52
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

03/16/2022 01:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____