
Midmark M11 - v1.0.5
Total Cycles: 996

Name: RITTER2
Sterilizer ID: V2313228
04/18/2022 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.1 F 0.0
2:00 178.0 F 0.4
4:00 186.6 F 2.4
6:00 206.6 F 7.2
8:00 225.9 F 12.4
10:00 231.7 F 8.6
12:00 242.1 F 11.1
14:00 243.8 F 12.9
16:00 244.1 F 12.4
18:00 257.3 F 19.5
20:00 268.9 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.7
1:00 272.6 F 29.5
2:00 272.0 F 28.7
3:00 272.5 F 29.5
4:00 272.3 F 28.9

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 207.6 F 0.0
10:00 213.7 F 0.0
15:00 219.3 F 0.0
20:00 223.7 F 0.0
25:00 227.9 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

04/18/2022 10:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____