

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4247

Name: RITTER 1  
Sterilizer ID: V2178198  
01/30/2024 10:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.5 F 0.0  
2:00 112.0 F 0.8  
4:00 114.3 F 2.5  
6:00 135.8 F 5.2  
8:00 161.7 F 8.8  
10:00 183.7 F 13.2  
12:00 213.5 F 5.0  
14:00 225.0 F 9.0  
16:00 236.6 F 13.3  
18:00 236.2 F 9.6  
20:00 239.1 F 9.8  
22:00 245.2 F 12.9  
24:00 256.5 F 18.8  
26:00 266.9 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.4 F 28.9  
2:00 271.2 F 28.2  
3:00 272.5 F 29.0  
4:00 271.4 F 28.2

Min 271.2 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.3 F 0.0  
5:00 218.6 F 0.0  
10:00 218.9 F 0.0  
15:00 224.5 F 0.0  
20:00 227.5 F 0.0  
25:00 231.2 F 0.0  
30:00 230.2 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 27:35  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:05:58  
CYCLE COMPLETE

01/30/2024 11:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_