

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001137  
8:56 23/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:32  
PRESSURIZING 3:34  
STERILIZING 4:49  
136.0 C 220kPa 4:49  
136.2 C 223kPa 5:19  
136.3 C 224kPa 5:49  
136.4 C 225kPa 6:19  
136.2 C 224kPa 6:49  
136.1 C 223kPa 7:19  
136.4 C 224kPa 7:49  
136.3 C 223kPa 8:19  
136.3 C 223kPa 8:19  
VENTING 8:19  
AIR DRYING 9:17  
CYCLE COMPLETE 20:17  
-----

Digital Signature #  
11BA40D6A047E327  
-----