

-----  
Midmark M11 - v1.0.5  
Total Cycles: 328

Name: RITTER2  
Sterilizer ID: V2313228  
09/02/2021 06:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.5 F 0.0  
2:00 80.4 F 0.0  
4:00 122.8 F 0.0  
6:00 138.0 F 0.1  
8:00 145.2 F 1.5  
10:00 173.7 F 5.1  
12:00 194.0 F 8.9  
14:00 211.0 F 12.7  
16:00 220.0 F 6.4  
18:00 230.8 F 10.4  
20:00 234.6 F 8.7  
22:00 242.1 F 12.4  
24:00 240.1 F 10.9  
26:00 252.9 F 17.3  
28:00 263.7 F 23.6  
30:00 271.9 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 271.9 F 29.1  
2:00 272.4 F 29.5  
3:00 271.6 F 29.6  
4:00 271.3 F 28.9  
  
Min 271.0 F 28.9  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.1 F 0.0  
5:00 216.7 F 0.0  
10:00 192.7 F 0.0  
15:00 189.9 F 0.0  
20:00 191.1 F 0.0  
25:00 185.7 F 0.0  
30:00 192.0 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 30:06  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:08:25  
CYCLE COMPLETE

09/02/2021 08:07 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02