
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006942
13:07 13/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:06
STERILIZING 4:03
136.5 C 221kPa 4:03
136.1 C 218kPa 4:33
136.2 C 222kPa 5:03
136.3 C 224kPa 5:33
136.4 C 225kPa 6:03
136.4 C 225kPa 6:33
136.4 C 221kPa 7:03
136.4 C 224kPa 7:33
136.4 C 224kPa 7:33
VENTING 7:33
AIR DRYING 8:32
CYCLE COMPLETE 8:37

Digital Signature #
6ABE17B0365F8C6C
