
Midmark M11 - v1.0.5
Total Cycles: 1541

Name: RITTER2
Sterilizer ID: V2313228
09/27/2022 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.8 F 0.0
2:00 183.4 F 3.7
4:00 200.3 F 9.3
6:00 227.5 F 7.3
8:00 235.5 F 10.7
10:00 233.8 F 8.8
12:00 238.6 F 9.8
14:00 249.2 F 15.5
16:00 262.2 F 22.8
18:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.3
2:00 271.9 F 29.7
3:00 272.3 F 29.2
4:00 271.8 F 29.8

Min 271.2 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 227.4 F 0.0
10:00 249.2 F 0.0
15:00 227.2 F 0.0
20:00 219.5 F 0.0
25:00 229.7 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 3:04
HEATING: 18:11
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

09/27/2022 10:04 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____