
Midmark M11 - v1.0.0
Total Cycles: 4813

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2024 08:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 83.1 F 0.0
4:00 85.4 F 0.1
6:00 118.1 F 0.3
8:00 149.6 F 0.5
10:00 174.2 F 0.8
12:00 199.3 F 2.3
14:00 212.4 F 5.7
16:00 225.2 F 9.0
18:00 234.5 F 12.8
20:00 233.4 F 8.0
22:00 244.2 F 12.9
24:00 240.1 F 10.3
26:00 235.7 F 7.9
28:00 249.9 F 14.9
30:00 261.4 F 21.4
32:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.8
2:00 272.1 F 28.2
3:00 271.6 F 28.1
4:00 272.7 F 28.8

Min 271.5 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 213.7 F 0.0
10:00 197.5 F 0.0
15:00 218.5 F 0.0
20:00 189.9 F 0.0
25:00 184.7 F 0.0
30:00 181.9 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 32:39
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:43
CYCLE COMPLETE

08/14/2024 09:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____