

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4634

Name: RITTER 1  
Sterilizer ID: V2178198  
06/13/2024 08:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 93.6 F 0.0  
4:00 87.3 F 0.1  
6:00 126.3 F 0.2  
8:00 154.4 F 1.4  
10:00 179.3 F 4.7  
12:00 200.5 F 8.1  
14:00 213.5 F 11.5  
16:00 219.9 F 4.0  
18:00 227.7 F 7.9  
20:00 237.3 F 11.9  
22:00 234.2 F 8.1  
24:00 246.6 F 13.2  
26:00 243.3 F 11.8  
28:00 255.1 F 17.8  
30:00 265.5 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.6 F 28.7  
2:00 271.4 F 28.2  
3:00 272.6 F 28.6  
4:00 271.3 F 28.1

Min 271.2 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.9 F 0.0  
5:00 229.1 F 0.0  
10:00 231.9 F 0.0  
15:00 231.9 F 0.0  
20:00 223.0 F 0.0  
25:00 226.4 F 0.0  
30:00 230.0 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:50  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:09:59  
CYCLE COMPLETE

06/13/2024 09:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_