
Midmark M11 - v1.0.5
Total Cycles: 3457

Name: RITTER2
Sterilizer ID: V2313228
03/15/2024 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.8 F 0.0
2:00 142.9 F 1.0
4:00 152.7 F 3.3
6:00 174.8 F 6.8
8:00 192.9 F 10.9
10:00 213.0 F 4.9
12:00 221.9 F 8.7
14:00 233.4 F 13.3
16:00 236.8 F 10.4
18:00 236.2 F 8.9
20:00 248.0 F 14.7
22:00 259.2 F 20.8
24:00 269.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.9 F 29.8
2:00 271.3 F 29.4
3:00 272.6 F 29.8
4:00 271.2 F 29.2

Min 271.2 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.6 F 0.0
5:00 200.3 F 0.0
10:00 196.9 F 0.0
15:00 196.3 F 0.0
20:00 192.8 F 0.0
25:00 189.7 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:53
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:30
CYCLE COMPLETE

03/15/2024 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____