
Midmark M11 - v1.0.0
Total Cycles: 4514

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2024 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 72.9 F 0.0
4:00 89.7 F 0.5
6:00 109.8 F 2.6
8:00 142.0 F 5.4
10:00 173.3 F 8.8
12:00 187.5 F 12.6
14:00 204.6 F 4.0
16:00 216.4 F 7.4
18:00 227.6 F 11.1
20:00 227.8 F 6.4
22:00 239.1 F 11.4
24:00 235.8 F 8.7
26:00 248.9 F 14.9
28:00 259.8 F 20.9
30:00 270.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.4 F 29.2
2:00 272.0 F 28.5
3:00 272.4 F 29.1
4:00 272.4 F 28.8

Min 271.5 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 204.1 F 0.0
10:00 219.1 F 0.0
15:00 219.4 F 0.0
20:00 222.1 F 0.0
25:00 224.5 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:48
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:08:59
CYCLE COMPLETE

05/02/2024 09:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____