
Midmark M11 - v1.0.5
Total Cycles: 2532

Name: RITTER2
Sterilizer ID: V2313228
06/19/2023 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.7 F 0.0
2:00 191.9 F 4.0
4:00 206.0 F 9.2
6:00 225.3 F 8.0
8:00 234.1 F 11.0
10:00 234.3 F 8.9
12:00 238.9 F 9.9
14:00 248.6 F 14.9
16:00 260.9 F 21.8
18:00 272.0 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.0 F 28.9
2:00 272.5 F 29.4
3:00 271.6 F 29.1
4:00 272.1 F 29.1

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 217.8 F 0.0
10:00 227.4 F 0.0
15:00 231.7 F 0.0
20:00 229.3 F 0.0
25:00 229.8 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 18:27
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:60
CYCLE COMPLETE

06/19/2023 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____