
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004704
13:14 14/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:57
STERILIZING 2:54
136.5 C 221kPa 2:54
135.9 C 220kPa 3:24
136.1 C 221kPa 3:54
136.2 C 223kPa 4:24
136.3 C 222kPa 4:54
136.2 C 219kPa 5:24
136.2 C 220kPa 5:54
136.4 C 224kPa 6:24
136.4 C 224kPa 6:24
VENTING 6:24
AIR DRYING 7:24
CYCLE COMPLETE 12:24

Digital Signature #
03B3504174593C9A
