
Midmark M11 - v1.0.5
Total Cycles: 1007

Name: RITTER2
Sterilizer ID: V2313228
04/19/2022 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.3 F 0.0
2:00 195.4 F 2.6
4:00 210.5 F 8.6
6:00 225.3 F 9.1
8:00 235.2 F 11.5
10:00 236.6 F 10.1
12:00 237.4 F 9.5
14:00 251.8 F 17.0
16:00 264.4 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.4 F 29.7
2:00 271.9 F 29.8
3:00 272.2 F 29.5
4:00 271.9 F 29.8

Min 271.5 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 221.8 F 0.0
10:00 224.5 F 0.0
15:00 229.8 F 0.0
20:00 228.8 F 0.0
25:00 228.5 F 0.0
30:00 226.9 F 0.0

Drying Complete

FILLING: 2:43
HEATING: 17:49
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:46
CYCLE COMPLETE

04/19/2022 03:18 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____