
Midmark M11 - v1.0.5
Total Cycles: 1201

Name: RITTER2
Sterilizer ID: V2313228
06/20/2022 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.0 F 0.0
2:00 169.8 F 2.9
4:00 173.3 F 6.5
6:00 191.4 F 11.1
8:00 217.8 F 5.0
10:00 228.8 F 10.1
12:00 230.9 F 8.4
14:00 243.3 F 12.7
16:00 241.8 F 11.4
18:00 254.4 F 18.0
20:00 265.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.6
2:00 272.2 F 29.1
3:00 271.2 F 28.7
4:00 272.0 F 29.6

Min 271.2 F 28.6
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 214.2 F 0.0
10:00 218.7 F 0.0
15:00 218.8 F 0.0
20:00 219.6 F 0.0
25:00 219.5 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

06/20/2022 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____