
Midmark M11 - v1.0.0
Total Cycles: 2491

Name: RITTER 1
Sterilizer ID: V2178198
10/26/2022 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.9 F 0.0
2:00 174.1 F 3.6
4:00 183.8 F 7.9
6:00 205.4 F 12.8
8:00 216.3 F 6.9
10:00 230.0 F 12.1
12:00 234.1 F 8.7
14:00 239.8 F 10.5
16:00 245.1 F 13.0
18:00 257.3 F 19.6
20:00 268.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.3 F 29.5
2:00 271.8 F 28.7
3:00 272.3 F 29.5
4:00 271.8 F 28.8

Min 271.4 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 229.1 F 0.0
10:00 224.8 F 0.0
15:00 240.5 F 0.0
20:00 233.8 F 0.0
25:00 219.7 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

10/26/2022 11:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____