

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 008404  
11:52 11/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:48  
PRESSURIZING 1:53  
STERILIZING 2:46  
136.4 C 222kPa 2:46  
136.4 C 225kPa 3:17  
136.1 C 219kPa 3:47  
135.9 C 220kPa 4:17  
136.4 C 223kPa 4:47  
136.4 C 220kPa 5:17  
136.1 C 222kPa 5:47  
136.4 C 221kPa 6:17  
136.4 C 221kPa 6:17  
VENTING 6:17  
AIR DRYING 7:04  
CYCLE COMPLETE 37:04

-----  
Digital Signature #  
C485069C769E7539  
-----