

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4007

Name: RITTER2  
Sterilizer ID: V2313228  
10/07/2024 02:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.0 F 0.0  
2:00 122.2 F 0.2  
4:00 147.9 F 2.8  
6:00 172.0 F 6.2  
8:00 194.4 F 9.8  
10:00 208.9 F 9.7  
12:00 220.0 F 7.3  
14:00 232.2 F 11.6  
16:00 232.1 F 8.2  
18:00 243.6 F 13.0  
20:00 243.1 F 12.3  
22:00 254.5 F 18.3  
24:00 264.9 F 24.6  
26:00 271.2 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.5 F 29.8  
2:00 271.5 F 29.1  
3:00 271.8 F 29.8  
4:00 272.2 F 29.6

Min 271.2 F 29.0  
Max 272.5 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
5:00 187.0 F 0.0  
10:00 200.7 F 0.0  
15:00 194.8 F 0.0  
20:00 186.9 F 0.0  
25:00 181.9 F 0.0  
30:00 192.9 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 26:01  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:04:48  
CYCLE COMPLETE

10/07/2024 03:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_