
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000725
9:31 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:22
STERILIZING 3:26
136.0 C 229kPa 3:26
136.4 C 231kPa 3:57
136.0 C 230kPa 4:27
136.3 C 232kPa 4:57
136.4 C 233kPa 5:27
136.2 C 231kPa 5:57
136.1 C 230kPa 6:27
136.3 C 233kPa 6:57
136.3 C 233kPa 6:57
VENTING 6:57
AIR DRYING 7:48
CYCLE COMPLETE 7:55

Digital Signature #
1F6C6FBB151EC113
