
Midmark M11 - v1.0.5
Total Cycles: 2552

Name: RITTER2
Sterilizer ID: V2313228
06/22/2023 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.3 F 0.0
2:00 189.4 F 3.8
4:00 201.5 F 9.0
6:00 220.9 F 8.6
8:00 231.5 F 10.2
10:00 233.0 F 8.1
12:00 242.7 F 12.1
14:00 245.1 F 13.1
16:00 257.4 F 19.8
18:00 268.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.7
2:00 272.0 F 29.1
3:00 271.2 F 28.9
4:00 272.4 F 29.7

Min 271.2 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 215.1 F 0.0
10:00 236.8 F 0.0
15:00 232.2 F 0.0
20:00 218.3 F 0.0
25:00 211.6 F 0.0
30:00 209.4 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

06/22/2023 11:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____