
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002211
15:05 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:41
STERILIZING 4:05
136.5 C 226kPa 4:05
136.4 C 225kPa 4:35
136.3 C 223kPa 5:05
136.4 C 223kPa 5:35
136.3 C 225kPa 6:05
136.4 C 225kPa 6:35
136.4 C 224kPa 7:05
136.1 C 223kPa 7:35
136.1 C 223kPa 7:35
VENTING 7:35
AIR DRYING 8:29
CYCLE COMPLETE 11:45

Digital Signature #
3AF22CC8094C16D1
