
Midmark M11 - v1.0.0
Total Cycles: 5270

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2025 04:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.1 F 0.0
2:00 156.3 F 3.1
4:00 176.1 F 6.8
6:00 216.5 F 10.9
8:00 213.6 F 3.9
10:00 226.8 F 8.2
12:00 238.0 F 12.6
14:00 235.3 F 8.8
16:00 241.7 F 11.2
18:00 244.4 F 12.4
20:00 255.9 F 18.4
22:00 266.3 F 24.7
24:00 272.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.7 F 29.1
2:00 271.7 F 28.8
3:00 272.8 F 29.1
4:00 271.5 F 28.7

Min 271.5 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 199.1 F 0.0
10:00 190.0 F 0.0
15:00 188.2 F 0.0
20:00 188.8 F 0.0
25:00 199.9 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:22
CYCLE COMPLETE

01/23/2025 06:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____