
Midmark M11 - v1.0.5
Total Cycles: 3284

Name: RITTER2
Sterilizer ID: V2313228
01/10/2024 02:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.5 F 0.0
2:00 130.6 F 0.8
4:00 156.6 F 3.5
6:00 180.7 F 7.2
8:00 198.9 F 11.4
10:00 219.6 F 5.4
12:00 225.0 F 10.0
14:00 232.1 F 9.0
16:00 240.6 F 12.6
18:00 240.5 F 11.7
20:00 253.3 F 18.2
22:00 264.9 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.5
1:00 272.4 F 30.1
2:00 271.2 F 29.4
3:00 272.1 F 30.3
4:00 272.5 F 30.2

Min 271.0 F 29.4
Max 273.0 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.4 F 0.0
5:00 225.6 F 0.0
10:00 214.3 F 0.0
15:00 183.7 F 0.0
20:00 180.6 F 0.0
25:00 178.2 F 0.0
30:00 176.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:52
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

01/10/2024 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____