
Midmark M11 - v1.0.5
Total Cycles: 2374

Name: RITTER2
Sterilizer ID: V2313228
05/16/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 217.3 F 0.0
2:00 209.1 F 3.9
4:00 220.3 F 9.1
6:00 225.9 F 8.2
8:00 237.5 F 11.4
10:00 235.2 F 9.4
12:00 236.7 F 9.3
14:00 249.6 F 15.5
16:00 262.0 F 22.5
18:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.4 F 29.4
2:00 271.3 F 29.0
3:00 272.4 F 29.5
4:00 271.6 F 28.9

Min 271.2 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 225.4 F 0.0
10:00 236.7 F 0.0
15:00 229.1 F 0.0
20:00 222.6 F 0.0
25:00 222.2 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:20
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

05/16/2023 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____