
Midmark M11 - v1.0.5
Total Cycles: 4444

Name: RITTER2
Sterilizer ID: V2313228
03/21/2025 03:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 163.6 F 2.5
4:00 181.2 F 7.0
6:00 206.0 F 12.4
8:00 219.6 F 6.5
10:00 232.7 F 11.8
12:00 234.6 F 9.4
14:00 237.9 F 9.6
16:00 249.3 F 15.5
18:00 262.0 F 22.9
20:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 273.0 F 29.9
2:00 271.9 F 29.1
3:00 272.6 F 30.1
4:00 272.7 F 29.6

Min 271.6 F 29.0
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 219.8 F 0.0
10:00 230.2 F 0.0
15:00 230.0 F 0.0
20:00 227.0 F 0.0
25:00 225.5 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:15
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

03/21/2025 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____