
Midmark M11 - v1.0.0
Total Cycles: 3841

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2023 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.8 F 0.0
2:00 189.3 F 2.9
4:00 208.0 F 7.4
6:00 215.5 F 12.3
8:00 223.2 F 6.8
10:00 233.8 F 11.7
12:00 234.6 F 8.7
14:00 241.4 F 10.8
16:00 246.2 F 13.3
18:00 258.0 F 19.6
20:00 268.6 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 28.8
2:00 271.3 F 28.2
3:00 272.5 F 28.8
4:00 271.2 F 28.1

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 227.9 F 0.0
10:00 236.4 F 0.0
15:00 233.6 F 0.0
20:00 228.5 F 0.0
25:00 228.8 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

09/29/2023 12:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____