

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3405

Name: RITTER 1  
Sterilizer ID: V2178198  
06/19/2023 11:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
2:00 186.7 F 3.1  
4:00 206.4 F 8.2  
6:00 224.1 F 12.1  
8:00 232.2 F 9.3  
10:00 240.0 F 9.6  
12:00 242.8 F 12.7  
14:00 241.7 F 11.3  
16:00 254.7 F 17.8  
18:00 265.9 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.7 F 29.3  
2:00 271.9 F 28.6  
3:00 271.6 F 28.8  
4:00 272.5 F 29.2  
  
Min 271.3 F 28.3  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 236.4 F 0.0  
10:00 221.6 F 0.0  
15:00 228.4 F 0.0  
20:00 227.3 F 0.0  
25:00 228.5 F 0.0  
30:00 229.0 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 19:44  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

06/19/2023 12:57 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_