

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2960

Name: RITTER2  
Sterilizer ID: V2313228  
10/02/2023 02:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.8 F 0.0  
2:00 183.5 F 4.0  
4:00 208.9 F 9.9  
6:00 220.9 F 5.9  
8:00 235.8 F 11.6  
10:00 237.1 F 9.9  
12:00 236.9 F 9.0  
14:00 252.4 F 16.7  
16:00 265.0 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 271.9 F 28.5  
2:00 272.2 F 28.8  
3:00 272.7 F 29.2  
4:00 272.3 F 29.3  
  
Min 271.6 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
5:00 213.8 F 0.0  
10:00 226.2 F 0.0  
15:00 228.9 F 0.0  
20:00 227.8 F 0.0  
25:00 226.8 F 0.0  
30:00 224.1 F 0.0  
  
Drying Complete

FILLING: 2:37  
HEATING: 17:39  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:29  
CYCLE COMPLETE

10/02/2023 03:51 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_