
Midmark M11 - v1.0.0
Total Cycles: 4398

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2024 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.8 F 0.0
2:00 120.4 F 1.3
4:00 133.7 F 3.9
6:00 171.0 F 7.3
8:00 217.9 F 12.0
10:00 219.0 F 4.1
12:00 224.6 F 8.3
14:00 234.2 F 12.8
16:00 234.7 F 8.9
18:00 241.2 F 11.1
20:00 244.6 F 12.3
22:00 256.2 F 18.4
24:00 266.9 F 24.8
26:00 272.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 271.6 F 28.5
2:00 271.6 F 28.6
3:00 271.6 F 28.5
4:00 271.7 F 28.4

Min 271.5 F 28.3
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.5 F 0.0
5:00 251.0 F 0.0
10:00 237.6 F 0.0
15:00 225.8 F 0.0
20:00 221.6 F 0.0
25:00 217.0 F 0.0
30:00 214.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 26:09
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:04:42
CYCLE COMPLETE

03/22/2024 02:06 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____