
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001723
12:12 27/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:44
STERILIZING 3:56
136.4 C 225kPa 3:56
136.4 C 225kPa 4:26
136.4 C 225kPa 4:56
136.2 C 224kPa 5:26
136.2 C 224kPa 5:56
136.4 C 225kPa 6:26
136.3 C 224kPa 6:56
136.3 C 222kPa 7:26
136.3 C 222kPa 7:26
VENTING 7:26
AIR DRYING 8:18
CYCLE COMPLETE 11:15

Digital Signature #
42C9DDC39EA1C832
