

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3419

Name: RITTER2  
Sterilizer ID: V2313228  
02/28/2024 03:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 88.5 F 0.0  
2:00 88.4 F 0.0  
4:00 101.2 F 0.4  
6:00 124.9 F 2.0  
8:00 155.2 F 5.0  
10:00 179.0 F 8.2  
12:00 196.4 F 11.6  
14:00 210.5 F 4.4  
16:00 217.8 F 8.5  
18:00 229.7 F 12.7  
20:00 233.5 F 9.8  
22:00 236.4 F 9.9  
24:00 245.5 F 13.8  
26:00 256.5 F 19.7  
28:00 266.6 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.5  
1:00 272.6 F 30.3  
2:00 271.9 F 29.7  
3:00 271.6 F 30.0  
4:00 272.9 F 30.6

Min 271.0 F 29.4  
Max 273.3 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 221.6 F 0.0  
10:00 213.8 F 0.0  
15:00 189.3 F 0.0  
20:00 180.6 F 0.0  
25:00 178.7 F 0.0  
30:00 177.3 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 29:43  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:08:21  
CYCLE COMPLETE

02/28/2024 04:21 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_