
Midmark M11 - v1.0.0
Total Cycles: 2735

Name: RITTER 1
Sterilizer ID: V2178198
01/04/2023 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 169.6 F 0.0
4:00 196.2 F 1.9
6:00 212.4 F 6.2
8:00 223.6 F 11.2
10:00 228.1 F 6.9
12:00 239.7 F 12.4
14:00 238.5 F 10.2
16:00 236.1 F 8.3
18:00 251.1 F 15.7
20:00 263.1 F 22.7
22:00 271.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.7
2:00 271.6 F 28.4
3:00 272.3 F 28.7
4:00 271.4 F 28.3

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 224.3 F 0.0
10:00 242.4 F 0.0
15:00 235.2 F 0.0
20:00 223.7 F 0.0
25:00 221.1 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

01/04/2023 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____