
Midmark M11 - v1.0.0
Total Cycles: 2829

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2023 05:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.3 F 0.0
2:00 176.1 F 5.0
4:00 194.7 F 9.9
6:00 215.7 F 5.3
8:00 227.5 F 9.2
10:00 235.4 F 9.4
12:00 241.9 F 12.0
14:00 240.4 F 10.4
16:00 254.1 F 17.4
18:00 266.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.7
2:00 271.5 F 28.3
3:00 272.6 F 28.8
4:00 271.7 F 28.1

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 223.2 F 0.0
10:00 230.6 F 0.0
15:00 229.5 F 0.0
20:00 229.3 F 0.0
25:00 227.9 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

01/30/2023 06:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____