
Midmark M11 - v1.0.0
Total Cycles: 4235

Name: RITTER 1
Sterilizer ID: V2178198
01/26/2024 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.3 F 0.0
2:00 153.5 F 1.5
4:00 158.3 F 4.5
6:00 180.2 F 7.9
8:00 198.0 F 11.7
10:00 209.8 F 3.5
12:00 219.7 F 7.5
14:00 231.0 F 11.6
16:00 230.4 F 7.4
18:00 242.1 F 12.4
20:00 239.9 F 10.4
22:00 251.3 F 16.0
24:00 262.1 F 22.1
26:00 271.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 29.0
2:00 271.6 F 28.3
3:00 271.7 F 28.7
4:00 272.3 F 28.9

Min 271.1 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 207.2 F 0.0
10:00 207.9 F 0.0
15:00 209.8 F 0.0
20:00 214.1 F 0.0
25:00 216.3 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:36
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:05:21
CYCLE COMPLETE

01/26/2024 11:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____